



THROUGH E-MAIL

H. P. GOVERNMENT DENTAL COLLEGE AND HOSPITAL, SHIMLA-1

Ph. No. : 0177-2658838, 2652562, Fax No.: 0177-2651483

Website: www.hpgdcshimla.org

Email: hpgdcshimla@gmail.com

H.P. DENTAL
COLLEGE & HOSPITAL SHIMLA
RECEIVED

Ref. No. HFW (GDC)G(6)-3/99-IV- **1470**

Dated: **12 MAR 2025**

To

1. M/S CRB International, SCO No. 78, First Floor, Sector-39-D, Chandigarh-160036
2. M/S Unicorn Denmart Ltd. 3, LSC, Rajender Nagar, New Delhi-110060
3. M/S Himachal Dental Co., New Bypass Road, Duganri, Hamirpur, HP -177001
4. M/S Confident Dental Equipment Ltd. A-28, Mayapuri Phase-II, Opp. New Era Public School, New Delhi 110064
5. M/S Nobal Biocare, Kuldeep Nagar Nanhera, Ambala Cantt, Haryana-133004
6. M/S Compact Dental Point, New Bypass Road, Hamirpur, HP -177001
7. College Website/Notice Board, HP,GDC&H, Shimla


Subject:- Quotation for supply Dental Physio Dispenser.

Sir/Madam,

It is to inform you that sealed quotations are hereby invited by the undersigned for the supply of Dental Physio Dispenser for Implant to be used in this institution as per specifications enclosed. Quotations should reach in the office of the undersigned on or before **26-03-2025 by 1.00 PM**. Thereafter no quotation will be entertained. Quotations should be neatly typed. These quotations will be opened on the same day i.e. on **26-03-2025 by 2:30PM** in the presence of the bidders or their authorized representative's who-so-ever likes to be present. The rates will be approved of L-1 bidder after demonstration. The required quantity may vary. The detail of the items is as under:-

Sr. No.	Description of Equipment	Quantity	Specifications	Name of Brand/ Company	Rate per unit	GST %	Total
1.	Dental Physio Dispenser	02	As per list enclosed				

The specifications of Dental Physio Dispenser and terms & condition of tender document are attached herewith Annexure-A, B & C.


PRINCIPAL,
H.P. GOVERNMENT DENTAL COLLEGE
& HOSPITAL SHIMLA
And Hospital, Shimla

Enclosed:- Terms and Condition.

H. P. GOVERNMENT DENTAL COLLEGE AND HOSPITAL, SHIMLA-1

TERM & CONDITIONS

"ANNEXURE- A"

- 1) The quotations are to be submitted in a sealed cover containing two separately one envelope (**Envelope-A**) super-scribed as Technical Bid and the offer/price on the prescribed Performa in the second envelope (**Envelope-B**) super-scribed as Financial bids (i.e. rate of Physio Dispenser) and both the envelopes A & B shall be put in single envelope i.e.

i) **Technical Bid Envelop:-** This envelop shall contain having following documents:-


1. Technical specifications.
2. Undertaking to acceptance of terms & conditions of the tender Annexure "A" both duly signed along with attested copies of required document.
3. Power of Attorney.
4. Copy to PAN and GST No.
5. Undertaking on stamp paper that rate quoted by them are not higher than rate quoted to any Govt. Deptt. during last one year.
6. Undertaking that they will not involve in cartelization for the bid.
7. Demonstration/sampling (Physical or Online) as per mutual understanding will also be part of technical evaluation.

ii) **Financial Bid Envelop:-** This envelop shall contain "Annexure-B" of rates only. The financial bid of the technically rejected bidders shall not be opened.

- 2) Technical bid envelop will be opened first and financial bid shall be opened of the technically qualified bids only.
- 3) Incomplete quotation/conditional tender will not be accepted.
- 4) Rate should be FOR basis.
- 5) Minimum warranty period shall be 03 year or standard warranty by OEM whichever is higher followed by 05 year CMC of product. Rate of CMC should not be accepted beyond 05% (five) of the product cost.
- 6) Rate of **Dental Physio Dispenser** should be valid of one year i.e. up-to 31-3-2026.
- 7) Performance security of ₹10,000 /-(Ten Thousand only) shall be deposited by the successful bidder and agreement shall be executed with the institution.
- 8) The financial bid evaluated on cumulative quoted rate of product CMC for 05 years after 03 years warranty period taken into account for the L-1 Selection on the prescribed format with financial bid:-

Name of the Equipment	Basic Price without GST	1 st year CMC	2 nd year CMC	3 rd year CMC	4 th year CMC	5 th year CMC	TOTAL

- 9) Successful bidder will provide to produce a certificate that they are the authorized distributors /manufacturer and the rates have been charged in accordance with the manufacturer's price as being charged from other Govt. institution/Hospital.


PRINCIPAL,
H.P. GOVT. DENTAL COLLEGE
& HOSPITAL, Shimla

"ANNEXURE- C"

TECHNICAL SPECIFICATIONS OF THE DENTAL PHYSIO DISPENSER

Power Supply	AC 100-240, V 50/60 Hz
Max. Pump Output	75 mL/min
Programs / Implant Systems	8 programs
Torque	5-70 NCM
Motor Speed	200-40,000 per min
Light Power (Optic Micromotor)	Over 32,000 LUX
Foot Control Functions	Coolant Solution Flow Volume Button, PRG (Program) Button Forward/Reverse Button, Speed Control
Gear Ratio	20:1 Reduction.
Hand piece	Fiber optic motor.
Water Spray	External and internal cooling.
Body Material	Stainless steel.
<ul style="list-style-type: none">• The unit should LCD display with micro-processor controlled and brushless electronic single motor• Motor and its cooling system should be safe for implantology and should have versatility under sterile condition in order to avoid infection and tissue damage• Should have a feather touch panel for easy operation• Should have single function keys with haptic and audible feedback• Unit should have built in torque controller with safety for overload protection• Should have selectable torque level upto 70 Ncm• Should have programmable speed control with facility for varying the speed with foot pedal• Unit should give alarm when in reverse mode• The unit should have built in irrigation pump for cooling system with 10 step adjustment for flow rate• Unit should be quoted with all essential accessories and hand pieces including straight handpiece, 20:1 c contrangled reduction handpiece• Hands free operation via foot control unit• Powered with advanced autoclavable handpiece caliberation• Optic brushless micro motor with lowest noise and vibration level• User friendly simple operation	



Principal
H.P.Govt. Dental College
And Hospital, Shimla